



Declaration of Conformity

I, Nigel Leigh, of Commtest Instruments Ltd., declare that the following equipment complies with EC standard IEC/EN 61326:2002 "Electrical equipment for measurement, control and laboratory use - EMC requirement" (including IEC 61000-4-2:1995 to IEC 61000-4-5:1995, IEC 61000-4-6:1996 and IEC 61000-4-11:1994):

Vibration Analyzer - vb7

This equipment complies with the essential requirements of the EMC Directive 89/336EEC of the European Community.

Authorized Representative:

Nigel Leigh

A handwritten signature in blue ink that reads "Nigel Leigh".

Date:

A handwritten date in blue ink that reads "13-July-2007".

Parkside Laboratories



GLOBAL TESTING, CALIBRATION & CERTIFICATION SERVICES

◆ ELECTRICAL ◆ EMC ◆ ENVIRONMENTAL ◆

Laboratory Test Report

*This report may not be reproduced except in full.
It contains no corrections or erasures.*

Report No.: 11667-00
Date Issued: 06 December 2006

SUBJECT: Portable Vibration Analyser
Model: vbMerino
Rating: 7.5 V d.c, 400 mA

REQUESTED BY: Commtest Instruments Limited
28b Moorhouse Avenue
P.O. Box 9297
Christchurch 8001
NEW ZEALAND

INSTRUCTIONS: Test for compliance with IEC 61326:2002 "Electrical equipment for measurement, control and laboratory use - EMC requirements".

CONTENTS: General
Test Specification
Date of test
Description
Results: IEC 61326:2002
Appendix A: CISPR 16-2-1:2003
Appendix B: CISPR 16-2-3:2003
Appendix C: IEC 61000-4-2:1995
Appendix D: IEC 61000-4-3:1995
Appendix E: IEC 61000-4-4:1995
Appendix F: IEC 61000-4-5:1995
Appendix G: IEC 61000-4-6:1996
Appendix H: IEC 61000-4-11:1994
Scans: - Radiated Emissions 30 MHz – 200 MHz
- Radiated Emissions 200 MHz – 1 GHz
- Conducted Emissions 150 kHz – 30 MHz

SUMMARY: All test results in this report in relation to the Portable Vibration Analyser Model vbMerino confirmed that the specimen complied with the relevant provisions of IEC 61326:2002.

APPROVED BY:

Manuel Shimasaki
IANZ Signatory

TESTED BY:

Vinesh Chand
EMC Compliance Engineer

PREPARED BY:

Amy Dunn
Admin Assistant